

References:

1. Chapman MD, Aalberse RC, Brown MJ, Platts-Mills TA. Monoclonal antibodies to the major feline allergen Fel d 1. 11. Single step affinity purification of Fel d 1, N-terminal sequence analysis, and development of a sensitive two-site immunoassay to assess Fel d 1 exposure. *J Immunol* 1988; 140(3):812-818.
2. de Blay F, Chapman MD, Platts-Mills TA. Airborne cat allergen (Fel d 1). Environmental control with the cat in situ. *Am Rev Respir Dis* 1991; 143:1334-39.
3. Ingram JM, Sporik R, Rose G, Honsinger R, Chapman MD, Platts-Mills TA. Quantitative assessment of exposure to dog (Can f 1) and cat (Fel d 1) allergens: relation to sensitization and asthma among children living in Los Alamos, New Mexico. *J Allergy Clin Immunol* 1995; 96(4):449-456.
4. Patchett K, Lewis S, Crane J, Fitzharris P. Cat Allergen (Fel d 1) levels on school children's clothing and in primary school classrooms in Wellington, New Zealand. *J Allergy Clin Immunol* 1997; 100(6 Pt 1):755-759
5. Platts-Mills TA, Vervloet D, Thomas WR, Aalberse RC, Chapman MD. Indoor allergens and asthma: report of the Third International Workshop. *J Allergy Clin Immunol* 1997; 100(6 Pt 1):S2-24.
6. Platts-Mills TA, Vaughan J, Squillace S, Woodfolk J, Sporik R. Sensitisation, asthma, and a modified Th2 response in children exposed to cat allergen: a population-based cross-sectional study. *Lancet* 2001; 357(9258):752-756.
7. Arbes SJ, Jr., Cohn RD, Yin M, Muilenberg ML, Friedman W, Zeldin DC. Dog allergen (Can f 1) and cat allergen (Fel d 1) in US homes: results from the National Survey of Lead and Allergens in Housing. *J Allergy Clin Immunol* 2004; 114(1):111-117.
8. Heinrich J, Bedada GB, Zock JP, Chinn S, Norback D, Oliveri M et al. Cat allergen level: its determinants and relationships to specific IgE to cat across European centers. *J Allergy Clin Immunol* 2006; 118(3):674-681
9. Jarvis D, Zock JP, Heinrich J, Svanes C, Verlato G, Olivieri M et al. Cat and dust mite allergen levels, specific IgG and IgG(4), and respiratory symptoms in adults. *J Allergy Clin Immunol*. 2007 Mar; 119(3):697-704.



www.inbio.com

Indoor Biotechnologies, Inc.
1216 Harris Street,
Charlottesville,
VA 22903
United States

Tel: (434) 984-2304
Fax: (434) 984-2709
E-mail: mail@inbio.com

Indoor Biotechnologies Ltd.
Cardiff Medicentre
Heath Park
Cardiff CF14 4UJ
Wales

Tel: +44 (0) 2920 68 2115
Fax: +44 (0) 2920 68 2044
E-mail: info@indoorbiotech.co.uk




Fel d 1 ELISA kit (6F9/3E4)

Product Code: EL-FD1

Lot Number: xxxxx

Sample Curve:



Content:

Vial 1 (red top) 100 μ L
Monoclonal antibody 6F9
Concentration: 1mg/ml in PBS

Vial 2 (white top) 400 μ L
Universal Allergen Standard
Concentration: 1000ng/ml Fel d 1

Vial 3 (brown) 100 μ L
Biotinylated Monoclonal antibody 3E4
Dilute: 1:1000 for use

Storage: The ELISA kit should be stored at 4°C

For research and commercial use in vitro: not for human in vivo or therapeutic use.

Certificate of Analysis

Monoclonal Antibody: 6F9 (clone 6F9 A4 H1)
Immunogen: Fel d 1
Isotype: Mouse IgG1
Specificity: Binds to an epitope on cat *Felis domesticus* allergen, Fel d 1, Chain 1.
Purification: Produced in tissue culture and purified by affinity chromatography using recombinant Protein G. Single heavy and light chain bands on SDS-PAGE.
Concentration: 1.0mg/ml in phosphate buffered saline, pH 7.4. Based on A280 for IgG (1.42=1mg/ml) 0.22µm filtered, preservative free.

Lot Number: xxxxx

Monoclonal Antibody: 3E4 (clone 3E4 C4 C10)
Immunogen: Fel d 1
Isotype: Mouse IgG1
Specificity: Binds to an epitope on cat *Felis domesticus* allergen, Fel d 1, Chain 1.
Purification: Produced in ascites and purified by ammonium sulphate precipitation and HPLC using recombinant Protein G. Single heavy and light chain bands on SDS-PAGE.
Biotinylation: Biotinylated and titrated for use in ELISA at 1/1000 dilution. Prepared in 1% BSA/50% glycerol/PBS, pH 7.4, 0.22µm filtered, preservative free.

Lot Number: xxxxx

Allergen Standard: Universal Allergen Standard
Composition: A formulation of eight purified natural allergens prepared in 1% BSA/50% glycerol PBS, pH 7.4.

Lot Number: xxxxx

Concentration:

Universal Allergen Standard	Protein Measurement	Concentration (ng/ml)
Der p 1	Amino-acid analysis	2500
Der f 1	Amino-acid analysis	2500
Der p 2	Amino-acid analysis	1000
Fel d 1	Amino-acid analysis	1000
Can f 1	Amino-acid analysis	2500
Rat n 1	Amino-acid analysis	1000
Mus m 1	Amino-acid analysis	250
Bla g 2	Amino-acid analysis	2500

ELISA Protocol for Fel d 1.

- Coat polystyrene microtiter plates (NUNC Maxisorp Cert. NUNC catalog # 439454) with 100µl mAb 6F9 at 10µl/10ml, i.e. 1/1000 dilution of stock, in 50mM carbonate-bicarbonate buffer, pH 9.6, incubate overnight at 4°C.
- Wash wells 3x with PBS-0.05% Tween 20, pH 7.4 (PBS-T). Incubate for 30 min. at room temperature with 100µl/well of 1% BSA, PBS-T. Wash 3x with PBS-T.
- Use doubling dilutions of the Universal Allergen Standard to make a control curve ranging from 100 - 0.2ng/ml Fel d 1: Pipette 20µl UAS into 180µl 1% BSA, PBS-T into wells A1 and B1 on the ELISA plate. Mix well and transfer 100µl across the plate into 100µl 1% BSA, PBS-T diluent to make 10 serial doubling dilutions. Wells A11, B11 and A12, B12 should contain only 1% BSA, PBS-T as blanks.
- Add 100µl of diluted allergen samples and incubate for 1 hour at room temperature. House dust extracts for Fel d 1 analysis are routinely diluted two-fold from 1/10-1/80. Other sample types, like air filter extracts and allergen extracts, may require different dilutions.
- Wash wells 3x with PBS-T and add 100µl diluted biotinylated anti-Fel d 1 mAb 3E4. The antibody solution contains 50% glycerol and should be diluted 1/1000 in 1%BSA, PBS-T. Incubate for 1 hour at room temperature.
- Wash wells 3x with PBS-T and add 100µl diluted Streptavidin-Peroxidase (Sigma S5512, 0.25mg reconstituted in 1ml distilled water). The reconstituted Streptavidin should be diluted 1/1000 in 1% BSA, PBS-T. Incubate for 30 minutes at room temperature.
- Wash wells 3x and develop the assays by adding 100µl 1mM ABTS in 70mM citrate phosphate buffer, pH 4.2 and 1/1000 dilution of H₂O₂. Read the plate when the absorbance at 405nm reaches 2.0-2.4.

Notes:

Universal Allergen Standard is recommended for immunoassay calibration purposes only. Not recommended for in-vitro antibody measurements, T cell studies, immunization purposes, or other uses.

Buffer recipes, storage conditions and a list of frequently asked questions can be found under "Protocols" on our web site: www.inbio.com.

For research and commercial use in vitro: not for human in vivo or therapeutic use.